

## Laundering and Care

- 1. Remove all panels and/ or any metal frames or components prior to cleaning.
- 2. This brace can be hand washed in cold or lukewarm water with mild soap or detergent and air dried
- 3. Do not use any substance containing Chlorine

### Warning

- 1. Carefully read this application instruction and warning prior to fitting braces, failure to properly position the brace and fasten each strap will compromise performance and support.
- 2. This support will Not prevent or reduce all injuries. Proper rehabilitation and activity modification are essential while wearing the brace.
- 3. Federal Law restricts the sale of this Device to anyone other than a Licensed Health Care Provider.
- 4. Settings of valgus/varus controls should be undertaken by medical professionals only

## Caution

- 1. Single patient use only
- 2. Do not over-tighten closures
- 3. If irritation, pain or discomfort occurs while wearing this brace, discontinue use and immediately consult your medical professional.

#### FUROPEAN AUTHORIZED REPRESENTATIVE



Wellkang Ltd Suite B, 29 Harley Street, LONDON, W1G 9QR, U.K. www.CE-Marking.EU

Councial Directive concerning Medical Device MANUFACTURED BY

## Dicarre LLC

info@dicarre.com

## Toll Free-1-844-DICARRE

500 N. Rainbow Blvd., Suite 300.

Las Vegas, NV 89107 www.dicarre.com




DK03 Three Panel Knee Immobilizer

DK05

Single Panel Knee Immobilizer


User Guide Application Instruction



# Airmesh Pocket Sling

- Undo all the contact closure straps
- 2. Position your knee on the knee immobilizer
- 3. The removable panels (DKO3)may be adjusted laterally for correct sizing of thigh and calf circumference
- 4. Wrap and fasten the straps closest to the centre of the knee
- 5. Then proceed to fasten the upper thigh and calf straps.
- 6. Adjust the straps to a snug fit.

DK03 Dicarre Three Panel Knee Immobilizer



DK05 Dicarre Single Panel Knee Immobilizer

